



Consumer and  
Corporate Affairs Canada

Consommation  
et Corporations Canada

Legal Metrology  
Branch

Direction de la  
Métrologie légale

Holland Avenue  
Ottawa, Ontario  
K1A 0C9

Avenue Holland  
Ottawa, Ontario  
K1A 0C9

Your file    Votre référence

Our file    Notre référence

Telephone: (613)952-0640

File N°: 06635  
Project: AP-GL-88-0060

Canadian Meter Co.  
3037 Derry Road, West  
Milton, Ontario  
L9T 2X6  
Attn.: Mr. G.F. Wilson

1989-04-03

Dear Mr. Wilson,

This is in response to your recent notification given in accordance with the Electricity and Gas Inspection Regulations, Section 13, regarding devices approved under Notices of Approval G-205, G-233 and G-238. It is noted that these devices are to be modified as follows:

- The incorporation of a modified software package, (identified as "software-pkg 1.2 (52699D001)"), expanding the display capability, on command, of information retained in the Electronic Volume Corrector.
- This modification does not, in any way, influence the calculation of corrected volume by the E.V.C.

I am advised by the Gas Metrology Laboratory that their evaluation of the proposed modifications indicates that they may be regarded as immaterial in relation to the approval provisions of the Act. Accordingly, your notification is acknowledged and the modifications are accepted as being included within the above-noted approval. All recipients of the approval notice are being informed by copy of this letter.

Yours truly,

W.R. Virtue, Chief  
Legal Metrology Laboratories

cc: Approvals Mailing List

Canada