



Industry Canada  
Legal Metrology Branch

Industrie Canada  
Direction de la métrologie légale

Holland Avenue  
Ottawa, Ontario  
K1A 0C9

Avenue Holland  
Ottawa (Ontario)  
K1A 0C9

Telephone: (613) 952-0639

Téléphone: (613) 952-0639

Project: AP-GL-94-0023  
Date: 1994-07-28

Mr. Paul Honchar  
Equimeter Inc.  
805 Liberty Blvd.  
P.O. Box 528  
DuBois, Penn., USA  
15801

Dear Mr. Honchar,

This is in response to your recent notification given in accordance with the Electricity and Gas Inspection Regulations, Section 13, regarding devices approved under Notices of Approval G-204 Rev. 1 to Rev. 5 inclusive, G-252, G-252 Rev. 1, G-253 and G-253 Rev. 1. It is noted that these devices are to be modified as follows:

- New firmware version identified as Version 3.2. The new firmware changes the instrument prompt from "Rockwell" to "Equimeter".

I am advised by the Gas Laboratory that their evaluation of the proposed modifications indicates that they may be regarded as immaterial in relation to the approval provisions of the Act. Accordingly, your notification is acknowledged and the modifications are accepted as being included within the above-noted approval. All recipients of the approval notice are being informed by copy of this letter.

Yours truly,

H.L. Fraser, Director  
Electricity and Gas

cc: Approvals Mailing List



Industry Canada  
Legal Metrology Branch

Industrie Canada  
Direction de la métrologie légale

Holland Avenue  
Ottawa, Ontario  
K1A 0C9

Avenue Holland  
Ottawa (Ontario)  
K1A 0C9

Telephone: (613) 952-0639

Téléphone: (613) 952-0639

Projet: AP-GL-94-0023

Date: 1994-07-28

M. Paul Honchar  
Equimeter Inc.  
805 Boul. Liberty  
DuBois, Pennsylvanie  
É-U 15801

Monsieur,

Nous accusons réception de l'avis de modification que vous nous avez envoyé récemment conformément à l'article 13. du Règlement sur l'inspection de l'électricité et du gaz, au sujet des appareils approuvés en vertu des avis d'approbation G-204 Rév. 1 à Rév. 5 inclusivement, G-252, G-252 Rév. 1, G-253 et G-253 Rév. 1. Il est à noter que ces appareils doivent être modifiés comme suit:

- Nouvelle version de la microprogrammation désignée Version 3.2. Dans cette nouvelle microprogrammation, le message de l'instrument indique "Equimeter" au lieu de "Rockwell".

Après avoir étudié les modifications proposées, le Laboratoire de gaz me signale que ces modifications ne contreviennent en rien aux dispositions de la Loi. En conséquence, les modifications sont acceptées et intégrées de facto dans les avis d'approbation susmentionnés. Tous ceux qui ont reçu l'avis d'approbation sont avisés de cette décision par une copie de la présente lettre.

Je vous prie d'agréer mes salutations distinguées.

H.L. Fraser, Chef  
Électricité et gaz

c.c.: Liste d'envoi des approbations