



CANADA

DEPARTMENT OF TRADE AND COMMERCE
STANDARDS DIVISION

SD-GA-3

OTTAWA, October 20, 1948.

NOTIFICATION OF TYPE APPROVAL

The apparatus specified and illustrated herein has been duly approved under the provisions of the appropriate Act, and may be admitted to verification in Canada.

Authority:- The Gas Inspection Act, Chapter 82, R.S. 1927.

Apparatus Approved:- "Emco" Size $4\frac{1}{2}$ Gas Meter, manufactured by the Pittsburgh Equitable Meter Company, Pittsburgh, Pa., U.S.A., and distributed in Canada by Peacock Bros., Limited, Montreal, P.Q.


Application:- Measurement in distribution service of manufactured or natural gas, or mixtures thereof.


Rating of Apparatus:- 6000 cubic feet per hour.

Limiting Conditions:- Approval granted, provided that the words "per revolution" or "per rev" be marked or stamped under the words "Cubic Feet" on the index.

Description:- The "Emco" Size $4\frac{1}{2}$ is a pressed steel gas meter and, except for somewhat smaller size in proportion to the reduced capacity, is identical in construction with the Size 5, previously approved. The index carries seven indication dials, with the lowest reading dial registering 10 cubic feet per revolution.

Testing:- The standard tests for a gas meter of this rating will apply.


L. V. Hunt,
Chief Engineer


A. F. Gill,
Director of Standards