



TRADE AND COMMERCE
CANADA

SD-CA.29

STANDARDS DIVISION

OTTAWA, December 13, 1952.

TYPE APPROVAL

CANADIAN METER COMPANY TYPE "5M-3"
POSITIVE DISPLACEMENT GAS METER


The apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 82, R.S. 1927, as amended, and may be admitted to verification in Canada.

Apparatus Approved: Type "5M-3" Tinned Steelcase Positive Displacement Gas Meter, manufactured by the Canadian Meter Company Limited, Hamilton, Ontario.

Rating of Apparatus:

Side Tubes $\frac{1}{4}$ " or 1"
Badged Capacity 270 cu.ft./hr.(air)
Differential Pressure
at badged capacity 0.5" water gauge
Capacity per rev. 0.166 cubic foot.

Description: This meter is identical to the type "5M" except for larger valve seats and covers which allow for the additional gas capacity.


R. W. MacLean,
Director,
Standards Division.



E. F. Power,
Assistant Director (E&G),
Standards Division.

Ref: A-268A