

TRADE AND COMMERCE
CANADA

SD-GA.37-a

STANDARDS DIVISION

OTTAWA, October 29, 1953.

MODIFICATION OF TYPE APPROVALPARKINSON & COWAN TYPE "D-4" POSITIVE DISPLACEMENT GAS METER

The modification to the apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 82, R.S. 1927, as amended, and the modified apparatus may be admitted to verification in Canada.


Apparatus Modified: Type "D-4" Tinned Steelcase Positive Displacement Gas Meter, manufactured by Parkinson & Cowan (Gas Meters) Limited, London, England, and assembled in Canada by Hamer Meter Company, Toronto, Ontario.

Rating of Apparatus: (As previously approved in Circular SD-GA.37)

Modification: The modification consists of permitting the use of this meter with $1\frac{1}{2}$ " connections in addition to the $1\frac{1}{4}$ " connections previously approved.

E. F. Power

E. F. Power,
Assistant Director (E&G),
Standards Division.



R. W. MacLean,
Director,
Standards Division.