

TRADE AND COMMERCE  
CANADA

SD-GA.40

## STANDARDS DIVISION

OTTAWA, July 31, 1953.

TYPE APPROVALAMERICAN METER COMPANY 'STRAIGHT READING' GAS METER INDEX

The apparatus specified and illustrated herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 82, R.S. 1927, as amended, and may be admitted to verification in Canada.

Apparatus Approved: 'Straight Reading' Gas Meter Index, manufactured by the American Meter Company Inc., Erie, Pa., U.S.A., and distributed in Canada by the Canadian Meter Company Limited, Hamilton, Ontario.

Rating of Apparatus: Approved for use with any approved type and rating of gas meter manufactured or distributed by the Canadian Meter Company.

Description: This is a cyclometer type index. The main face of the dial has white letters on a black background, and the moving figures behind the windows are also white figures on black background.

The test dials in the illustration show 1/2 cu.ft. and 2 cu.ft. Other registers, depending on the size of the meter, could have test dials showing 1/4 cu.ft. and 1 cu.ft.; 1 cu.ft. and 5 cu.ft.; etc.

E. F. Power,  
Assistant Director (E&G),  
Standards Division.

R. W. MacLean,  
Director,  
Standards Division.

Ref: A-322