

TRADE AND COMMERCE
CANADA

SD-GA.44

STANDARDS DIVISION

OTTAWA, October 26, 1953.

MODIFICATIONS OF TYPE APPROVALPARKINSON & COWAN TYPES "D-1", "D-1A" AND "D-4"
POSITIVE DISPLACEMENT GAS METERS

The modifications to the apparatus specified herein have been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 82, R.S. 1927, as amended, and the modified apparatus may be admitted to verification in Canada.

Apparatus Modified: Types "D-1", "D-1A" and "D-4" Positive Displacement Gas Meters, manufactured by Parkinson & Cowan (Gas Meters) Limited, London, England, and assembled in Canada by Hamer Meter Company, Toronto, Ontario.

Rating of Apparatus: As previously approved - see Circular Letters:-
SD-GA.4, dated June 8, 1950
SD-GA.14, dated April 28, 1952
SD-GA.37, dated July 3, 1953.

Modifications: The modifications which are approved herewith are two in number:-

- 1) The pointers of the test dial and the 100-cu.ft. revolution dial are to be coloured "RED" in future; the pointers of the other dials will be "BLACK" as at present.
- 2) The external greasing device for the register spindle will not be fitted on future production.

E. F. Power

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean,
Director,
Standards Division.