

TRADE AND COMMERCE
CANADA

SD-GA.46

STANDARDS DIVISION

OTTAWA, December 3, 1953.

MODIFICATION OF TYPE APPROVALAMERICAN METER COMPANY TYPE "2B/50" GAS METER

The modification to the apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 82, R.S. 1927, as amended, and the modified apparatus may be admitted to verification in Canada.

Apparatus Modified: Type "2B/50" Aluminum Die-Cast Gas Meter, manufactured by the American Meter Company, Albany, N.Y., and distributed in Canada by the Canadian Meter Company Limited, Hamilton, Ontario.

Rating of Apparatus: 50 cubic feet per hour.

Modification: The type "2B/50" aluminum die-cast positive displacement gas meter, granted approval on Circular SD-GA.6 of August 16, 1950, was originally designed for use with liquid petroleum gases. For this reason it was fitted with a top which was tapped to accommodate $\frac{3}{8}$ " or $\frac{1}{2}$ " standard pipe for the Inlet and Outlet on the left and right side respectively.

Since there is now a demand for using the meter for measuring small quantities of manufactured or natural gas, the manufacturer proposes to fit some meters with what he terms the "Utility" top. This top has standard $\frac{1}{4}$ " 10-light connections on it.

Approval is granted herewith for the use of this "Utility" top.

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. Maclean,
Director,
Standards Division.