


## (Mechanism of Integrating Orifice Meter)



description of the operation is given on page 17 of American Meter Bulletin EG-2 "Orifice Meters".

The integrating mechanism is operated by a Telechron electric motor which has sufficient power to drive it continuously at constant speed. The synchronous motor also drives the meter chart through appropriate gearing. Motors may be supplied 110- or 220-volt, 60-cycle; may be either explosion-proof or standard; and may be either one or two revolutions per minute.

*E. F. Power*

E. F. Power,  
Assistant Director (E&G),  
Standards Division.

*R. W. MacLean*  
R. W. MacLean,  
Director,  
Standards Division.

Ref: A-318A