



TRADE AND COMMERCE
CANADA

SD-GA.75

STANDARDS DIVISION

OTTAWA, September 1, 1955.

TYPE APPROVAL

J. BRAUN AND COMPANY
TYPE "TWO" POSITIVE DISPLACEMENT GAS METER

The apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Type "Two" Positive Displacement Gas Meter, manufactured by J. Braun and Company, Stuttgart, Germany, and distributed in Canada by Burson Industries, Winnipeg, Manitoba.

Rating of Apparatus:

Badged Capacity 210 cu.ft./hr.(air)
Differential Pressure at badged capacity 0.5" water gauge
Capacity per revolution175 cu.ft.
Connections 1" male

Application: Measurement in distribution service of manufactured, natural, and petroleum gases, or mixtures thereof.

Description: The type "Two" meter is essentially the same as the type "One" approved on SD-GA.71, the difference being a higher capacity instrument.

E. F. Power

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean
R. W. MacLean,
Director,
Standards Division.

Ref: A-437