



TRADE AND COMMERCE  
CANADA

STANDARDS DIVISION

OTTAWA, September 21, 1955.

TYPE APPROVAL

BEGWACO TYPE "BANTAM"  
POSITIVE DISPLACEMENT GAS METER

The apparatus specified and illustrated herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Type "Bantam" Positive Displacement Gas Meter, manufactured by Begwaco Meters (Canada) Limited, No.40 Highway, Chatham, Ontario.

Rating of Apparatus:

Badged Capacity ..... 225 cu.ft./hr.(air)  
Differential Pressure at badged capacity ..... 0.5" water gauge  
Capacity per revolution ..... 0.0625 cu.ft.  
Connections ..... 1" male

Application: Measurement in distribution service of manufactured, natural, and petroleum gases, or mixtures thereof.

Description: The type "Bantam" gas meter is a two-diaphragm type. The case is die-cast aluminum. The valves and seats are bakelite and move in a radial direction. Diaphragms are of leather, or synthetic. To adjust the meter for accuracy, adjustment gears are provided which can be exchanged when the meter is proved.

*E. F. Power*

E. F. Power,  
Assistant Director (ESG),  
Standards Division.

*R. W. MacLean*  
R. W. MacLean,  
Director,  
Standards Division.

Ref: A-489

