



TRADE AND COMMERCE
CANADA

SD-GA.79

STANDARDS DIVISION

OTTAWA, September 21, 1955.

TYPE APPROVAL

AMERICAN METER COMPANY TYPE "W-60"
POSITIVE DISPLACEMENT GAS METER

The apparatus specified and illustrated herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Type "W-60" Positive Displacement Gas Meter, manufactured by the American Meter Company and distributed in Canada by the Canadian Meter Company Limited, Hamilton, Ontario.

Rating of Apparatus:

Badged Capacity 60 cu.ft./hr.(air)
Differential Pressure at badged capacity 0.5" water gauge
Connections $\frac{1}{2}$ " female

Application: Measurement in distribution service of manufactured, natural, and petroleum gases, or mixtures thereof.

Description: This meter has a welded steel case; moulded duramic three-convolution diaphragms. The formed, solid-center partition and sturdy lower flag rod bearings assure proper alignment of flag rods and diaphragms.

Approval for this meter will apply to three types produced:-

- (1) A soldered type without hand hole;
- (2) A soldered type with hand hole;
- (3) A welded top with hand hole.

E. F. Power

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean
R. W. MacLean,
Director,
Standards Division.

Ref: A-457 & A-427