



TRADE AND COMMERCE  
CANADA

SD-GA.88

## STANDARDS DIVISION

OTTAWA, November 21, 1955.

TYPE APPROVAL

CANADIAN METER COMPANY TYPE "5T"  
POSITIVE DISPLACEMENT GAS METER

The apparatus specified and illustrated herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Type "5T" Positive Displacement Gas Meter, manufactured by the Canadian Meter Company Limited, Milton, Ontario.

## Rating of Apparatus:

Badged Capacity ..... 210 cu.ft./hr.(air)  
Differential Pressure at badged capacity .... 0.5" water gauge  
Capacity per revolution ..... .166 cu.ft.  
Connections .....  $\frac{3}{4}$ " and 1"

Application: Measurement in distribution service of manufactured, natural, and petroleum gases, or mixtures thereof.

Description: This meter is basically the same as the type "5M-3" approved on December 13, 1952 on Circular SD-GA.29 except that it is an "open top" meter, that is, the valves and mechanism above the table are not enclosed in a valve box. The meter is Tin Steelcase construction and the top is equipped with a hand hold for convenience of repair and adjustment.

E. F. Power,  
Assistant Director (E&G),  
Standards Division.

R. W. MacLean,  
Director,  
Standards Division.

Ref: A-502