

TRADE AND COMMERCE
CANADA

SD-GA.89

STANDARDS DIVISION

OTTAWA, November 21, 1955.

MODIFICATION OF TYPE APPROVALCANADIAN METER COMPANY TYPE "5B-225"
ALUMINUMCASE POSITIVE DISPLACEMENT GAS METER

The modification to the apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and the modified apparatus may be admitted to verification in Canada.

Apparatus Modified: Type "5B-225" Aluminumcase Positive Displacement Gas Meter, manufactured by the Canadian Meter Company Limited, Milton, Ontario.

Rating of Apparatus:

Badged Capacity 170 cu.ft./hr.(air)
 Differential Pressure at badged capacity 0.5" water gauge
 Capacity per revolution111 cu.ft.
 Connections (male) 1 $\frac{1}{4}$ "

Modification: The modification permits the use of a three-convolution diaphragm and a different back and front drilled with 10 holes. In other respects the meter is similar to meters approved in Circulars SD-GA.60 and SD-GA.54.

E. F. Power,
 Assistant Director (E&G),
 Standards Division.

R. W. Maclean,
 Director,
 Standards Division.

Ref: A-507