



TRADE AND COMMERCE  
CANADA

SD-GA.91

## STANDARDS DIVISION

OTTAWA, December 8, 1955.

MODIFICATION OF TYPE APPROVAL

SMITH TYPE "U.D.MK.1"  
POSITIVE DISPLACEMENT GAS METER

The modification to the apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and the modified apparatus may be admitted to verification in Canada.

Apparatus Modified: Type "U.D.MK.1" Die-Cast Aluminumcase Positive Displacement Gas Meter, manufactured by Smith Meters Ltd., London, England.

## Rating of Apparatus:

Badged Capacity ..... 200 cu.ft./hr. (air)  
Differential Pressure at badged capacity ..... 0.3" water gauge  
Capacity per revolution ..... 0.1 cubic foot  
Connections (male) ..... 1½"

Modification: The modification permits the use of a four-way tangent, which may be supplied as an optional feature. The meter is similar in construction to the meter approved under Circulars SD-GA.64 and SD-GA.74 of February 9, 1955 and August 29, 1955 respectively.

*E. F. Power*

E. F. Power,  
Assistant Director (E&G),  
Standards Division.

*R. W. MacLean*  
R. W. MacLean,  
Director,  
Standards Division.

Ref: A-477