



TRADE AND COMMERCE
CANADA

SD-GA.93

STANDARDS DIVISION

OTTAWA, December 15, 1955.

TYPE APPROVAL

THOMAS GLOVER TYPE "A.1"
POSITIVE DISPLACEMENT GAS METER

The apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chap.129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Type "A.1" Positive Displacement Gas Meter, manufactured by Thomas Glover & Co. Limited, Gothic Works, London, England.

Rating of Apparatus:
Badged Capacity 100 cu.ft./hr.(air)
Differential Pressure at badged capacity 0.5" water gauge
Capacity per revolution1 cu.ft.
Connections (male) $\frac{1}{4}$ "

Application: Measurement in distribution service of manufactured, natural, and petroleum gases, or mixtures thereof.

Description: This meter is identical to the type "A.1" approved on Circular SD-GA.92 of December 15, 1955 except for the badged capacity and connections.

E. F. Power

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean
R. W. MacLean,
Director,
Standards Division.

Ref: A-472