

TRADE AND COMMERCE
CANADA

SD-GA.94

STANDARDS DIVISION

OTTAWA, February 3, 1956.

MODIFICATION OF TYPE APPROVALPARKINSON & COWAN TYPES "D-1" AND "D-1A"
POSITIVE DISPLACEMENT GAS METERS

The modification to the apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S.1952, and the modified apparatus may be admitted to verification in Canada.

Apparatus Modified: Types "D-1" and "D-1A" Positive Displacement Gas Meters, manufactured by Parkinson & Cowan (Gas Meters) Limited, London, England, and assembled in Canada by Duragas Equipment Ltd., Toronto, Ontario.

Rating of Apparatus: As previously approved - see Circulars:
SD-GA.4 - dated June 8, 1950
SD-GA.14 - dated April 28, 1952
SD-GA.44 - dated October 26, 1953.

Modifications: The modification concerns the stuffing box on the index assembly. In the new assembly, the gland nut is fitted to the front of the stuffing box and is accessible for re-packing purposes, after the index has been removed, without disturbing the gear box.

The re-designed stuffing box is regarded as an improvement which will facilitate meter repairs.

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean,
Director,
Standards Division.

Ref: A-448