

TRADE AND COMMERCE  
CANADA

SD-CA.98

## STANDARDS DIVISION

OTTAWA, May 28, 1956.

TYPE APPROVALAMERICAN METER COMPANY SERIES "A-20" ORIFICE METER

The apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chap.129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Type Series "A-20" 'Westcott' Orifice Meter, manufactured by the American Meter Company Inc., and distributed in Canada by the Canadian Meter Company Ltd., Milton, Ontario.

Application: Measurement of gas in conjunction with standard orifice plates.

## Rating of Apparatus:

Differential Ranges ..... 50 & 100 (inches of water)  
Working Pressure ..... 1000 (p.s.i.)

Description: Standard construction includes stainless steel internal parts, over-range and reverse-flow check valves, and adjustable pulsation dampeners. The pressure springs, cases, clocks, charts and other necessary parts are interchangeable with the Series "A-70" and "A-88" instruments, approved on Circular SD-GA.38 of June 25, 1953.

E. F. Power,  
Assistant Director (E&G),  
Standards Division.

R. W. MacLean,  
Director,  
Standards Division.

Ref: A-564