



TRADE AND COMMERCE
CANADA

SD-GA.99

STANDARDS DIVISION

OTTAWA, June 26, 1956.

MODIFICATION OF TYPE APPROVAL

SPRAGUE TYPE "1A" 'ZEPHYR'
ALUMINUMCASE OR IRONCASE GAS METERS

The modification to the apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and the modified apparatus may be admitted to verification in Canada.

Apparatus Modified: Type "1A" Aluminumcase or Ironcase Positive Displacement Gas Meters, manufactured by the Sprague Meter Company, Bridgeport, Connecticut, U.S.A.

Rating of Apparatus:

Badged Capacity 140 cu.ft./hr.(air)
Differential Pressure at badged capacity 0.5" water gauge
Capacity per revolution 0.133 cu.ft.
Connections (male) 1"

Modification: The modification permits the use, when required, of front and back covers made of cold rolled steel on the type "1A" aluminumcase or ironcase meters to replace the aluminum and iron front and back covers now used. (See Circular SD-GA.72 of July 7, 1955 for previous approval.)

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean,
Director,
Standards Division.