

TRADE AND COMMERCE
CANADA

SD-GA.104

STANDARDS DIVISION

OTTAWA, September 21, 1956.

TYPE APPROVALCANADIAN METER COMPANY TYPE "5T"
POSITIVE DISPLACEMENT GAS METER

The apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Type "5T" Positive Displacement Gas Meter, manufactured by the Canadian Meter Company Limited, Milton, Ontario.

Rating of Apparatus:

Badged Capacity	190 cu.ft./hr.(air)
Differential Pressure at badged capacity ...	0.5" water gauge
Capacity per revolution166 cu.ft.
Connections (male)	$\frac{1}{2}$ "

Application: Measurement in distribution service of manufactured, natural, and petroleum gases, or mixtures thereof.

Description: This meter is identical to the type approved under Circular SD-GA.88, November 21, 1955. The reduced capacity of the meter approved herein is due to the use of $\frac{1}{2}$ " connections. When $\frac{3}{4}$ " or 1" connections are used, the previously-approved capacity of 210 cu.ft./hr.(air) applies.

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLear,
Director,
Standards Division.

Ref: A-593