



TRADE AND COMMERCE  
CANADA

SD-GA.105

## STANDARDS DIVISION

OTTAWA, October 1, 1956.

TYPE APPROVAL

ROCKWELL TYPE "210"  
POSITIVE DISPLACEMENT ALUMINUMCASE GAS METER

The apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Rockwell Type "210" Positive Displacement Gas Meter, manufactured by the Rockwell Manufacturing Company, Pittsburgh, Pa., and distributed in Canada by its Canadian subsidiary, the Rockwell Manufacturing Co. of Canada Ltd., Guelph, Ontario.

## Rating of Apparatus:

Badged Capacity .....	160 cu.ft./hr.(air)
Differential Pressure at badged capacity ....	0.5" water gauge
Capacity per revolution .....	.0625 cu.ft.
Connections .....	1½" male.

Application: Measurement in distribution service of manufactured, natural, and petroleum gases, or mixtures thereof.

Description: This meter is similar to the type "250" meter approved on Circular SD-GA.70, June 7, 1955. A change has been made to the index box, with a one-sealing screw arrangement now being used.

E. F. Power,  
Assistant Director (E&G),  
Standards Division.

R. W. MacLean,  
Director,  
Standards Division.

Ref: A-594