

TRADE AND COMMERCE
CANADA

SD-GA.106

STANDARDS DIVISION

OTTAWA, November 13, 1956.

MODIFICATION OF TYPE APPROVALROCKWELL POSITIVE DISPLACEMENT GAS METERS

The modification to the apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and the modified apparatus may be admitted to verification in Canada.

Apparatus Modified: All previously-approved types of Positive Displacement Gas Meters manufactured by the Rockwell Manufacturing Company of Canada Ltd., Guelph, Ontario, which have, as an integral part of the covers of the index boxes, a small drain hole in the under side of the index box cover.

Modification: The modification consists only in a change in location of the drain hole in the index box. The new location for this drain hole, in all cases, will still be on the under side of the index box but as near as possible to the left-hand corner.

E. F. Power

E. F. Power,
Assistant Director (TAG),
Standards Division.

R. W. MacLean
R. W. MacLean,
Director,
Standards Division.

Ref: A-263B