


 TRADE AND COMMERCE
 CANADA

SD-GA.107

STANDARDS DIVISION

OTTAWA, November 15, 1956.

TYPE APPROVALCHARTS

The charts specified herein have been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Charts Approved: The following charts, submitted by the Canadian Meter Company, Limited, Milton, Ontario:

<u>Identification Number</u>	<u>Size</u>	<u>Use</u>	<u>Range</u>	<u>Time</u>	<u>Scale</u>
M-100-H-2	10"	Orifice Meter	0-100 Press. 0-100 Diff.	24-hr.	Linear
M-1000-T4	10"	Orifice Meter	0-1000 Press. 0-20 Diff.	24-hr.	Linear

E. F. Power,
 Assistant Director (E&G),
 Standards Division.

R. W. MacLean,
 Director,
 Standards Division.

Ref: A-598