



TRADE AND COMMERCE  
CANADA

STANDARDS DIVISION

OTTAWA,..... February 18, 1957.

TYPE APPROVAL

ROCKWELL #3 CAST IRON CASE  
POSITIVE DISPLACEMENT GAS METER

The apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Rockwell #3 Cast Iron Case Positive Displacement Gas Meter, manufactured by the Rockwell Manufacturing Company, Pittsburgh, Pa., and distributed in Canada by its subsidiary, the Rockwell Manufacturing Co. of Canada Ltd., Guelph, Ontario.

Rating of Apparatus:

Badged Capacity ..... 900 cu.ft./hr.(air)  
Differential Pressure at badged capacity .... 0.5" water gauge  
Capacity per revolution ..... 1.11 cu.ft.  
Connections ..... 3" female.

Application: Measurement in distribution service of manufactured, natural, and petroleum gases, or mixtures thereof.

Description: The design and performance of this instrument is similar to the #2<sup>1</sup>/<sub>2</sub> previously approved.

*E. F. Power*

E. F. Power,  
Assistant Director (E&G),  
Standards Division.

*R. W. MacLean*

R. W. MacLean,  
Director,  
Standards Division.

Ref: A-628