

TRADE AND COMMERCE  
CANADA

SD-GA.120

## STANDARDS DIVISION

OTTAWA, May 9, 1957.

MODIFICATION OF TYPE APPROVALBEGWACO BANTAM "300"  
POSITIVE DISPLACEMENT GAS METER

The modification to the apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and the modified apparatus may be admitted to verification in Canada.

Apparatus Modified: Type Bantam "300" Positive Displacement Gas Meter, manufactured by Begwaco Meters (Canada) Limited, No.40 Highway, Chatham, Ontario.

Rating of Apparatus: As previously approved. See Circular SD-GA.78 of September 21, 1955.

Modification: The modification permits the use of a sealed inspection hole in the top of the meter, also an adjustable tangent.

(for) E. F. Poyer,  
Assistant Director (E&G),  
Standards Division.

R. W. MacLean,  
Director,  
Standards Division.

Ref: A-642