



TRADE AND COMMERCE  
CANADA

SD-GA,128

## STANDARDS DIVISION

OTTAWA, August 2, 1957.

TYPE APPROVAL

ROCKWELL NO. 750 ALUMINUM CASE  
POSITIVE DISPLACEMENT GAS METER

The apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Rockwell No. "750" Positive Displacement Gas Meter, manufactured by the Rockwell Manufacturing Company, Pittsburg, Pa. and distributed in Canada by its subsidiary, the Rockwell Manufacturing Company of Canada Ltd., Guelph, Ontario.

## Rating of Apparatus:

Badged Capacity .....	600 cu.ft./hr.(air)
Differential Pressure at badged capacity ....	0.5" water gauge
Capacity per revolution .....	0.385 cu.ft.
Connections .....	1-1/4" & 1-1/2" male

Application: Measurement in distribution service of manufactured, natural, and petroleum gases, or mixtures thereof.

Description: This meter is similar in all respects to the Type "800" previously approved, November 20, 1956, on Circular SD-CA.109. Maximum working is 15 lbs. p.s.i.

E. F. Power,  
Assistant Director (E&G),  
Standards Division.

R. W. MacLean,  
Director,  
Standards Division.

Ref: A-683