

TRADE AND COMMERCE
CANADA

STANDARDS DIVISION

SD-GA.145

OTTAWA, September 25, 1958.

MODIFICATION OF TYPE APPROVALAMERICAN METER COMPANY INTEGRATOR
FOR TYPES "A-20", "A-70" AND "A-88"
ORIFICE METERS

The modification to the apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and the modified apparatus may be admitted to verification in Canada.

Apparatus Modified: Single and Double Integrating Mechanisms for use with types "A-20", "A-70" and "A-88" 'Westcott' Orifice Meters, manufactured by the American Meter Company Inc., and distributed in Canada by the Canadian Meter Company Limited, Milton, Ontario.

Modification: The modification consists only of an addition of a slip spring on the integrator pointer for differential pressure indication. The integrating mechanisms were previously approved in Circular SD-GA.61, December 24, 1954.

E. F. Power
E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean
R. W. MacLean,
Director,
Standards Division.

Ref: A-779