



TRADE AND COMMERCE
CANADA

SD-GA.147

STANDARDS DIVISION

OTTAWA, November 20, 1958.

TYPE APPROVAL

CANADIAN METER COMPANY
TYPES "80B", "250B" AND "500B"
ALUMINUMCASE POSITIVE DISPLACEMENT GAS METERS

The apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Types "80B", "250B" and "500B" Aluminumcase Positive Displacement Gas Meters, manufactured by the Canadian Meter Company Limited, Milton, Ontario.

Rating of Apparatus:	"80B"	"250B"	"500B"
Badged Capacity (cu.ft./hr. air)	860	2310	3750
Working Pressure (lbs.)	20	75 & 100	75 & 100
Differential Pressure at badged capacity (ins. w.g.)	0.5	0.5	0.5
Capacity per revolution (cu.ft.)833	1.67	4.00

Application: Measurement in distribution service of manufactured, natural, and petroleum gases, or mixtures thereof.

Description: Except for the aluminum case, the meters approved herein are similar in all respects to the meters previously approved in Circular SD-GA.146 of October 30, 1958.

E. F. Power

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean
R. W. MacLean,
Director,
Standards Division.

Ref: A-780