



TRADE AND COMMERCE
CANADA

SD-GA.154

STANDARDS DIVISION

OTTAWA, March 4, 1959.

MODIFICATION OF TYPE APPROVAL

SMITH TYPE "U.D. Mk/1"
POSITIVE DISPLACEMENT GAS METER

The modification to the apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and the modified apparatus may be admitted to verification in Canada.

Apparatus Modified: Type "U.D. Mk/1" Die-Cast Aluminumcase Positive Displacement Gas Meter, manufactured by Smith Meters Ltd., London, England, and distributed in Canada by Smith Meters (Canada) Limited, Toronto 18, Ontario.

Rating of Apparatus:

Badged Capacity	250 cu.ft./hr.(air)
Differential Pressure at badged capacity	0.5" water gauge
Capacity per revolution	0.1 cubic foot
Connections (male)	1 1/4"

Modification: The modification consists only in a change of valve cover material to bakelite in place of tin-antimony used previously. The meter was approved in Circular SD-GA.143, July 10, 1958.

E. F. Power

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean
R. W. MacLean,
Director,
Standards Division.

Ref: A-339H