



TRADE AND COMMERCE  
CANADA

## STANDARDS DIVISION

OTTAWA, May 26, 1959.

TYPE APPROVAL

CANADIAN METER COMPANY TYPE "AL-800"  
ALUMINUMCASE POSITIVE DISPLACEMENT GAS METER

The apparatus specified and illustrated herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Type "AL-800" Aluminumcase Positive Displacement Gas Meter, manufactured and distributed in Canada by the Canadian Meter Company Limited, Milton, Ontario.

Rating of Apparatus:

Badged Capacity .....	650 cu.ft./hr.(air)
Differential Pressure at badged capacity .....	0.5" water gauge
Capacity per revolution .....	0.385 cu.ft.
* Working Pressure .....	20 p.s.i. cond 35 p.s.i.
* Connections (male, top) <i>pi. female</i> .....	2" 1/2"

Application: Measurement in distribution service of manufactured, natural and petroleum gases or mixtures thereof.

Description: The meter approved herein is of the conventional design with corrugated body and top cover castings for added strength.

*E. F. Power*  
E. F. Power,  
Assistant Director (E&G),  
Standards Division.

*R. W. MacLean*  
R. W. MacLean,  
Director,  
Standards Division.

Ref: A-815

\* See Circular Letter July 2/60 SF-85-40