

TRADE AND COMMERCE
CANADA

SD-GA.157

STANDARDS DIVISION

OTTAWA..... May 27, 1959.

TYPE APPROVALSUPERIOR TYPE "215"
ALUMINUMCASE POSITIVE DISPLACEMENT GAS METER

The apparatus specified and illustrated herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Type "215" Aluminumcase Positive Displacement Gas Meter, manufactured and distributed in Canada by Neptune Meters Limited, Toronto, Ontario.

Rating of Apparatus:

Badged Capacity 170 cu.ft./hr.(air)
 Differential Pressure at badged capacity 0.5" water gauge
 Capacity per revolution077 cu.ft.
 Connections (male, top) 1 $\frac{1}{4}$ " pipe

Application: Measurement in distribution service of manufactured, natural and petroleum gases, or mixtures thereof.

Description: This meter is similar in all respects to types "250" and "340" previously approved on July 3, 1958 under Circular SD-GA.142 except that the rated capacity is lower.

E. F. Power,
 Assistant Director (E&G),
 Standards Division.

R. W. MacLean,
 Director,
 Standards Division.

Ref: A-796