

TRADE AND COMMERCE
CANADA

STANDARDS DIVISION

OTTAWA..... June 30, 1959.

TYPE APPROVAL

CANADIAN METER COMPANY
TYPES "25B", "35B", "80B", "250B", AND "500B"
IRONCASE POSITIVE DISPLACEMENT GAS METERS

The apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R. S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Types "25B", "35B", "80B", "250B" and "500B" Ironcase Positive Displacement Gas Meters, manufactured by the Canadian Meter Company, Limited, Milton, Ontario.

Rating of Apparatus:	"25B"	"35B"	"80B"	"250B"	"500B"
Badged Capacity (cu.ft./hr. air) ..	310	550	860	2310	3750
Working Pressure (lbs.)	100 & 250	100	100 & 250	250	250
Differential Pressure at badged capacity (ins. w.g.)	0.5	0.5	0.5	0.5	0.5
Capacity per revolution (cu.ft.) ..	.250	.455	.833	1.67	4.00

Application: Measurement in distribution service of manufactured, natural and petroleum gases, or mixtures thereof.

Description: The meters approved herein are identical with those approved in Circular SD-GA.146 of October 30, 1958 except that they are of stronger construction to withstand the higher pressures approved.

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean,
Director,
Standards Division.

Ref: A-780