



TRADE AND COMMERCE
CANADA

STANDARDS DIVISION

OTTAWA July 8, 1959.

TYPE APPROVAL

ROCKWELL TYPE #415 POSITIVE DISPLACEMENT GAS METER

The apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Type #415 Positive Displacement Gas Meter, manufactured and distributed in Canada by the Rockwell Manufacturing Company of Canada, Ltd., Guelph, Ontario.

Rating of Apparatus:

Badged Capacity	320 cu.ft./hr.(air)
Differential Pressure at badged capacity	0.5" water gauge
Capacity per revolution154 cu.ft.
Connections	various
Maximum Working Pressure	25 p.s.i.*

* to be marked on nameplate.

Description: The meter approved herein is the same in all respects to the type #415, 10 p.s.i. maximum working pressure, approved in Circular SD-GA.59 of December 20, 1954, with the following exceptions:-
The body to cover is connected with 5/16" - 18 x 1 1/4" long, hex head steel bolts; and the hand hole plate screws are 5/16" - 18 x 5/8" long, hex head steel.

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean,
Director,
Standards Division.

Ref: A-379