

TRADE AND COMMERCE
CANADA

STANDARDS BRANCH

S-GA.175

OTTAWA, December 11, 1959.

TYPE APPROVALDANIEL SIMPLEX ORIFICE PLATE HOLDERS

The apparatus specified and illustrated herein has been duly approved by the Standards Branch under the provisions of the Gas Inspection Act, Chap. 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Simplex Orifice Plate Holders, manufactured by the Daniel Orifice Fitting Company, Los Angeles, California, U.S.A., and distributed in Canada by Process-Instrument Systems Limited, 44 Haas Road, Rexdale P.O., Toronto, Ontario.

Rating of Apparatus:

"Flangnek" Standard Flange Facing

<u>Cast Steel</u>		<u>Range</u>
2" to 6"		150 A.S.A. 230 psi Gauge C.W.P.
2" to 6"		300 A.S.A. 600 psi Gauge C.W.P.
2" to 6"		600 A.S.A. 1200 psi Gauge C.W.P.
2" to 6" *		900 A.S.A. 1800 psi Gauge C.W.P.
2" to 6" *	Carbon Steel	1500 A.S.A. 3000 psi Gauge C.W.P.
2" to 6" *	Carbon Moly Alloy Steel	2500 A.S.A. 5000 psi Gauge C.W.P.

"Flangnek" Ring-Joint Flange Facing

2" to 6"		600 A.S.A. 1440 psi Gauge C.W.P.
2" to 6" *		900 A.S.A. 2160 psi Gauge C.W.P.
2" to 6" *	Carbon Steel	1500 A.S.A. 3600 psi Gauge C.W.P.
2" to 6" *	Carbon Moly Alloy Steel	2500 A.S.A. 6000 psi Gauge C.W.P.

Welding Neck Ends

2" to 6"		600 A.S.A. 1440 psi Gauge C.W.P.
2" to 6" *		900 A.S.A. 2160 psi Gauge C.W.P.
2" to 6" *	Carbon Steel	1500 A.S.A. 3600 psi Gauge C.W.P.
2" to 6" *	Carbon Moly Alloy Steel	2500 A.S.A. 6000 psi Gauge C.W.P.

.....(Socket)²
(over)