

TRADE AND COMMERCE
CANADA

S-GA.199

STANDARDS BRANCH

OTTAWA, January 3, 1961.

TYPE APPROVALCANADIAN METER COMPANY
TYPE "AL-425" POSITIVE DISPLACEMENT GAS METER

The apparatus specified herein has been duly approved by the Standards Branch under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Type "AL-425" Aluminum Case Positive Displacement Gas Meter, manufactured and distributed in Canada by the Canadian Meter Company Limited, Milton, Ontario.

Rating of Apparatus:

Badged Capacity	340 cu.ft./hr.(air)
Working Pressure	10 p.s.i.
Differential Pressure at badged capacity	0.5" water gauge
Capacity per revolution	0.200 cu.ft.
Connections (male)	1 $\frac{1}{4}$ " and 1 $\frac{1}{2}$ "

Application: Measurement in distribution service of heating gases.

Description: The meter approved herein is identical to the type "AL-425" approved on Circular S-GA.190 except that the meter may be fitted with 1 $\frac{1}{2}$ " male connections or with 1 $\frac{1}{4}$ " connections as previously approved.

E. F. Power,
Chief, Electricity & Gas Division,
Standards Branch.

R. W. MacLean,
Director,
Standards Branch.

Ref: A-868