



TRADE AND COMMERCE
CANADA

STANDARDS BRANCH

OTTAWA..... May 26, 1961.

TYPE APPROVAL

CANADIAN METER COMPANY CYCLOMETER REGISTER

The apparatus specified herein has been duly approved by the Standards Branch under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Cyclometer-type Register, manufactured and distributed by the Canadian Meter Company Limited, Milton, Ontario.

Rating of Apparatus: Approved for use with any approved type and rating of gas meter manufactured and distributed by the Canadian Meter Company Limited.

Description: This cyclometer-type register has a maximum reading of 999,900 cubic feet and one test dial of 1 cubic foot per revolution. An identification number '267' is stamped on the face plate.

E. F. Power

E. F. Power,
Chief, Electricity & Gas Division,
Standards Branch.

R. W. MacLean
R. W. MacLean,
Director,
Standards Branch.