



TRADE AND COMMERCE
CANADA

STANDARDS BRANCH

OTTAWA July 4, 1962.

TYPE APPROVAL

ROCKWELL TYPE "K-03" COUNTER REGISTER

The apparatus specified and illustrated herein has been duly approved by the Standards Branch under the provisions of the Gas Inspection Act, Chap. 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Counter Register Type "K-03", manufactured by the Rockwell Manufacturing Company, Pittsburgh, Pa., and distributed in Canada by its subsidiary, the Rockwell Manufacturing Company of Canada Limited, Guelph, Ontario.

Rating of Apparatus:

Maximum register capacity 9,999,990 cu.ft.
Test dial capacity 10 cu.ft. per revolution.

Description: The register has a six-digit counter suitably geared to a test dial pointer. The round face of the register forms a clock-type test dial with smallest subdivision corresponding to 0.1 cubic foot.

The register proper is mounted on the index plate assembly which carries underneath an intermediate gearing. The ratio of this gearing is so chosen that with a given meter capacity per revolution a number of complete tangent revolutions would correspond to one (or more) complete revolutions of the test dial pointer.

This register may be used with any large capacity approved type and rating positive displacement gas meter manufactured by the Rockwell Manufacturing Company, Pittsburgh, Pa., U.S.A.

J.R. Gardiner

(for) E. F. Power,
Chief, Electricity and Gas Division,
Standards Branch.

R.W. Maclean
R. W. Maclean,
Director,
Standards Branch.

Ref: A-963