



Consumer and
Corporate Affairs Canada

Consommation
et Corporations Canada

Legal Metrology
Branch

Direction de la
Métrologie légale

Holland Avenue
Ottawa, Ontario
K1A 0C9

Avenue Holland
Ottawa, Ontario
K1A 0C9

Legal Metrology Laboratories
Telephone: (613) 952-0639

Your file Votre référence

Our file Notre référence

O6635-B5
AP-GL-87-0036
March 21, 1988
AML-G-39

ITT Barton Instruments
3840 11A Street N.E.
Calgary, Alberta
T2E 6M6
Attention: Mr. Andy Seeger

Dear Mr. Seeger:


This is in response to your letter dated February 16, 1988, regarding Notice of Approval S-GA-242 and S-GA-250 and proposing to modify the models 202A and 208 Flow Recorders approved therein by recasting the case and door to (1) include a built-in rain shield, and (2) include the name "ITT BARTON" as part of the cast. Minor changes have also been made to the door hinges, door stop bushing, door stop latch and design of the nameplate.

I am advised that, upon evaluation of the proposed modifications, the Gas Laboratory has concluded that they may be regarded as immaterial in relation to the approval provisions of the Electricity and Gas Inspection Act. Accordingly, pursuant to the authority delegated to me under the said Act:

1. your letter is hereby acknowledged as satisfying the notification requirements of section 13, Part III of the Electricity and Gas Inspection Regulations, SOR/86-131, 16 January, 1986, and
2. the proposed alteration is hereby accepted, in accordance with the aforementioned section, as being an immaterial one and hence included within the approval.

All recipients of Notices of Approval under the Electricity and Gas Inspection Act are being informed of this decision by copy of this letter.

Yours truly,


W.R. Virtue
Chief

cc: Mailing List (Gas Approvals)

Canada