



TRADE AND COMMERCE
CANADA

STANDARDS DIVISION

3-GA.284

OTTAWA February 4, 1964.

TYPE APPROVAL

CANADIAN METER COMPANY CHART DRIVE "MIDGET"

MODEL 24-3

The apparatus specified and illustrated herein has been duly approved by the Standards Branch under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Chart drive, Type "Midget", Model 24-3, manufactured by Rockwell Manufacturing Company for the Canadian Meter Company and distributed by the Canadian Meter Company Limited, Milton, Ontario.

Rating of Apparatus: Wind period 8 days
Rotation period of output shaft 24 hours
Rotation period of output shaft with turret attachment 7 days.

Description: This spring driven chart drive has the standard clock mechanism. A turret attachment converts it to a 7-day rotation chart drive. It has a base plate drilled for mounting in standard and orifice meter cases. The type designation, model number, period of rotation of the output shaft, wind period and the distributor's name are given on the nameplate.

This Approval covers the use of this chart drive mechanism in any suitable, approved, circular chart recording instrument.

(for) E. F. Power,
Chief, Electricity & Gas Division,
Standards Branch.

R. W. MacLean,
Director,
Standards Branch.

Ref: 3L-100-25