

S-GA. -0242





Consumer and  
Corporate Affairs Canada

Consommation  
et Corporations Canada

Legal Metrology  
Branch

Direction de la  
Métrologie légale

Holland Avenue  
Ottawa, Ontario  
K1A 0C9

Avenue Holland  
Ottawa, Ontario  
K1A 0C9

Legal Metrology Laboratories  
Telephone: (613) 952-0639

Your file    *Voire référence*

Our file    *Notre référence*

O6635-B5  
AP-GL-87-0036  
March 21, 1988  
AML-G-39

ITT Barton Instruments  
3840 11A Street N.E.  
Calgary, Alberta  
T2E 6M6  
Attention: Mr. Andy Seeger

Dear Mr. Seeger:

This is in response to your letter dated February 16, 1988, regarding Notice of Approval (S-GA-242) and S-GA-250 and proposing to modify the models 202A and 208 Flow Recorders approved therein by recasting the case and door to (1) include a built-in rain shield, and (2) include the name "ITT BARTON" as part of the cast. Minor changes have also been made to the door hinges, door stop bushing, door stop latch and design of the nameplate.

I am advised that, upon evaluation of the proposed modifications, the Gas Laboratory has concluded that they may be regarded as immaterial in relation to the approval provisions of the Electricity and Gas Inspection Act. Accordingly, pursuant to the authority delegated to me under the said Act:

1. your letter is hereby acknowledged as satisfying the notification requirements of section 13, Part III of the Electricity and Gas Inspection Regulations, SOR/86-131, 16 January, 1986, and
2. the proposed alteration is hereby accepted, in accordance with the aforementioned section, as being an immaterial one and hence included within the approval.

All recipients of Notices of Approval under the Electricity and Gas Inspection Act are being informed of this decision by copy of this letter.

Yours truly,

W.R. Virtue  
Chief

cc: Mailing List (Gas Approvals)

**Canada**

