



CANADA

DEPARTMENT OF TRADE AND COMMERCE
STANDARDS BRANCH

S-GA.277

OTTAWA, October 17, 1963.

Type Approval

Rockwell Model 23-3 Chart Drive

The apparatus specified and illustrated herein has been duly approved by the Standards Branch under the provisions of the Gas Inspection Act, Chap. 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Rockwell Model 23-3 chart drive manufactured and distributed in Canada by the Rockwell Manufacturing Company of Canada, Ltd., Guelph, Ontario.

Rating of Apparatus:

Wind Period 31 days
Period of revolution of output shaft 7 days

Description: This spring driven chart drive has the standard clock mechanism and is supplied with jewelled escapement. It has a base plate drilled for mounting in standard instrument and orifice meter cases. The model number, period of rotation of output shaft, wind period, and the maker's name are clearly and permanently marked on the timing device.

This approval covers the use of this chart drive mechanism in any suitable, approved circular chart recording instrument. It should be noted that Chart Drive 23-1 is an unjewelled mechanism and similar in appearance to Drive 23-3. This unit has not been approved at this date.

E. F. Power

E. F. Power,
Chief, Electricity and Gas Division,
Standards Branch.

R. W. MacLean
R. W. MacLean,
Director,
Standards Branch.

Ref: A-995

ROCKWELL MODEL 23-3 CHART DRIVE



