

Ottawa, June 15, 1965.

SPG-4


SPECIAL APPROVAL

Granted to: Saskatchewan Power Corporation,  
Regina, Saskatchewan

Subject: Chromatographic Determination  
of B.t.u. Value of Natural Gas

Special Approval is granted to the Saskatchewan Power Corporation for the use of continuous gas sampling devices and chromatographic analyses of natural gas in the determination of B.t.u. value, subject to periodic checks made by the Standards Branch.

This will involve the analysis and determination of the result of our laboratory sample gas of known B.t.u. value by the Saskatchewan Power Corporation, witnessed by the District Inspector of Electricity and Gas, Standards Branch. This comparison is to be made at least once every three months, the tolerance to be the same as required for the recording calorimeter.



(K. Cryer)  
Chief, Electricity and Gas Division,  
Standards Branch.

KC/mee

c.c. to: Mr. W. E. Fletcher, District Inspector of Electricity and Gas,  
Regina, Saskatchewan.

