

SE-85-95  
SI-100-932  
SI-100-SPG

(A)

Ottawa, December 30, 1965.

SPECIAL APPROVAL

Granted to: Saskatchewan Power Corporation,  
Regina, Saskatchewan

Subject: Canadian UGC Recording Gas Gravitometer  
with Bristol Metameter Transmitter

Special Approval has been granted by the Standards Branch to the Saskatchewan Power Corporation for the use of the Bristol Metameter Transmitter with the UGC Recording Gas Gravitometer.

The inter-action by the Metameter Transmitter results in somewhat wider-than-usual recording pen trace and the instrument is calibrated to indicate the correct specific gravity of the gas at the lowest value of the chart record.



(K. Cryer)  
Chief, Electricity and Gas Division,  
Standards Branch.

LKW/mee

c.c. to: Mr. W. E. Fletcher, District Inspector of Electricity and Gas,  
Regina, Saskatchewan.

