SE-85-81 SI-100-822 SI-100-SPG

Ottawa, March 21, 1966

SPECIAL APPROVAL

Granted to: The Alberta Gas Trunk Line Company Limited,

505 Second Street, S.W.,

Calgary, Alberta.

Subject:

Kimray Gas Gravitometer

Special Approval has been granted to The Alberta Gas
Trunk Line Company Limited for the use, in connection with billing
for gas, of the Kimray Gas Gravitometer, Serial No. 174108, range
0.35 to 1.15 sp.gr., located in their laboratory in Calgary.

The gravitometer is intended for measurement of the specific gravity of gas samples brought in to the laboratory in stainless steel cylinders from various meter stations.

The instrument approved herewith shall be subjected to usual re-verification using our AC-ME Specific Gravity Balance as the standard of comparison.

Kloyer

(K. Cryer)
Chief, Electricity and Gas Division,
Standards Branch.

LKW/mee

c.c. to: Frank Wheatley Equipment (Calgary) Limited, 528 Cleveland Crescent S.E., Calgary, Alberta

c.c. to: Mr. A. S. Park, District Inspector of Electricity and Gas, Calgary, Alberta