

SE-85-81
SL-100-822
SL-100-SPG

Ottawa, February 23, 1967.

SPECIAL APPROVAL

Granted to: The Alberta Gas Trunk Line Company Limited,
505 Second Street, S.W.,
Calgary, Alberta.

Subject: Two Only Kimray Gas Gravitometers,
Serial Numbers 174114 and 174115,
Instrument Ranges 0.35 to 1.15.

Special Approval has been granted by the Standards Branch to The Alberta Gas Trunk Line Company Limited for use in Canada for billing purposes, to the above-named apparatus.

These gravimeters are to be located in the Company's Service Centre in Calgary and used to establish the gravity of gas samples brought in to the laboratory in stainless steel cylinders from various meter stations. They may also be taken to the field to spot-check gravity on the gas being recorded on field gravimeters.

These instruments shall be subjected to the quarterly re-verification by the Standards Branch.



(R. Cryer)

Chief, Electricity and Gas Division,
Standards Branch.

AKB/mee

c.c. to: Mr. A. S. Park, District Inspector of Elec. & Gas, Calgary

c.c. to: Frank Wheatley Industries Ltd.,
528 Cleveland Crescent S.E., Calgary

(Attention: Mr. A. P. Varro, General Manager)

Standards Branch

Department of Trade and Commerce

Data Sheet

1. Gas Meter Installation:
Utility: _____
Customer: _____
Address: _____

Date: _____
District: _____

2. Meter

Serial No.: _____
Manufacturer: _____
Model: _____

Size: _____

Approval: _____

Press. Rating: _____

3. Auxiliary Devices:

Pressure: _____

Range: _____

Approval: _____

Temperature: _____

Range: _____

Approval: _____

Other: _____

Approval: _____

4. Operating Conditions:

Flow;	Min. _____	Max. _____	Avge. _____
Pressure;	Min. _____	Max. _____	Avge. _____
Temperature;	Min. _____	Max. _____	Avge. _____

5. Flow Characteristics:

6. Dimensions Sketch:

