

SPG-9

SE-85-44
SL-100-855
SL-100-SPG

Ottawa, September 25, 1967.

SPECIAL APPROVAL

Granted to: The Alberta Gas Trunk Line Co. Ltd.
505 Second Street S.W.,
P. O. Box 2535,
Calgary, Alberta.

Subject: Rootsmeter Serial No. 6705614

(Attention: Mr. C. T. McCall, Gas Measurement Supt.)

Special Approval is hereby granted by the Standards Branch for The Alberta Gas Trunk Line Co. Ltd. to use the following meter to supply the Trans-Canada Pipe Lines Limited at Verger Homestead (near Brooks, Alberta):

Model: 3M200 Rootsmeter, Serial No. 6705614
Average Flowing Conditions: 650 psig, 45°F
Auxiliary Devices: Canadian Meter Company
Base Pressure Index, 0-1200#
Volume, Pressure & Temperature Gauge

Unit: Standard Cubic Feet

(Data furnished by Roots-Connersville (Canada) Limited
- D. K. Fairbrother - September 14, 1967)

Initial inspection and verification will be carried out by the District Inspector as usual.

(W. J. S. Fraser)
Chief, Electricity and Gas Division,
Standards Branch.

AKB/mee

c.c. to: Roots-Connersville (Canada) Limited, 21 Coronet Road, Toronto 18
- (attention: Mr. D. K. Fairbrother, Sales Manager)

c.c. to: Mr. A. S. Park, Dist. Insp. of Electricity and Gas, Calgary - with
Data Sheet
and Sketch

