

SPG-11

SL-87-51
SL-100-322
SL-100-SPG

Ottawa, December 23, 1967

SPECIAL APPROVAL

Granted to: Trans Canada Pipe Lines Limited
150 Eglinton Avenue East,
Toronto 12, Ontario.

Subject: Kinray Gas Gravimeter
Serial Number 174122
Instrument Range 0.35 to 1.15 sp.gr.

Special Approval has been granted by the Standards Branch to Trans Canada Pipe Lines Limited for use of the above-named apparatus in connection with billing for gas sales.

This gravimeter is located at the Emerson Sales Meter Station, Emerson, Manitoba.

The instrument shall be subjected to quarterly re-verification by the Standards Branch.

W. J. S. Fraser

(W. J. S. Fraser)
Chief, Electricity and Gas Division,
Standards Branch.

AKB/ana

- c.c. to: Mr. N. H. Young, District Inspector of Electricity and Gas,
Winnipeg, Manitoba
- c.c. to: Mr. J. D. Duggan, District Technical Supervisor,
Emerson Sales Meter Station, Trans Canada Pipe Lines Ltd.,
Emerson, Manitoba - (through Mr. Young)

