

SPG-15

SE-85-41
SL-100-SPG

OTTAWA, November 15, 1968.

SPECIAL APPROVAL

For: Rockwell Transducer Type I, Serial #B-3019

Special approval is hereby granted for a Rockwell Transducer Type "I" serial #B-3019 having a Range of Pressure (psi) 0-1200, and a range of temperature -20°F to 140°F.

This approval is based on the Type Approval S-GA-255 of March 1963 which granted approval for these units up to 500 psig.

Plains Western Gas Company intends to utilize this auxiliary with an R-3 Rotoseal meter at German Manitoba to check gas received from Trans Canada Pipe Lines.

W. J. S. Fraser
(W.J.S. Fraser)

cc Mr. H. H. McEwen, D.I. of ESG, Toronto
Plains Western Gas (Mr. H. Young)
Rockwell Manufacturing Co. of Canada Ltd. (Mr. D. Kennedy)

