

*S Re-16*

LL-89-81  
LL-100-822  
SL-100-5PG

OTTAWA, March 7, 1969.

SPECIAL APPROVAL

Granted to: The Alberta Gas Trunk Line Co. Ltd.,  
505 Second Street S.E.,  
Calgary, Alberta.

Subject: Kimray Gas Gravimeters -  
Serial numbers 174163 and 174144  
Instrument Ranges 0.35 to 1.15 sp.gr.

(Attention) Mr. C. F. McCall, Measurement Supt.)

Special Approval has been granted by the Standards Branch  
to the Alberta Gas Trunk Line Company Limited for use of the above-  
named apparatus in connection with billing for gas sales.

These gravimeters may also be used in the field to spot-  
check gravity on the gas being recorded on field gravimeters.

These instruments shall be subjected to quarterly re-  
verification by the Standards Branch.

*(W. J. S. Fraser)*

W. J. S. Fraser  
Chief,  
Electricity and Gas Division,  
Standards Branch.

A.S./ad

c.c. Mr. A. D. Park, Dist. Insp. of Gas, Calgary, Alberta.

