

SPG 21  
SFC-21

62-85-1A-2  
61-1A-786

OTTAWA, April 17, 1969.

TEMPORARY SPECIAL APPROVAL

Granted to: Quebec Natural Gas Corporation,  
1717 Du Havre,  
Montreal, Quebec.

Subject: Elevated Pressure Systems with Pressure  
Compensated Index.

Attention: Mr. A. Ledoux,  
Quebec Natural Gas Corporation

Temporary Special Approval is hereby granted by the Standards Branch to Quebec Natural Gas Corporation for the use of Pressure Compensated Indexes on specified installations. The addresses of the installations and details of the instrumentation are as follows:

1. Donald Pac Co., 10 La Province Ave., Ville Cartier, P.Q.
  - Meter Type AL-425, Serial 12-149085 (CNC 51818)  
Compensated at 2.14 psig, Compensated at 60°F
  - Pressure Recorder Serial 94-149141, two pen
  - Regulator Canadian Meter Co. Reliance, Type 1813B  
Serial 100-149145, Inlet pressure 60 psig, Internal relief set at 3 psig.
  
2. Uniroyal Chemical Division, 4145 Fourth Ave., Ville D'Anjou, P.Q.
  - Meter Type AL-425 Serial 12-149086 (CNC 51817)  
Compensated at 2.14 psig, Compensated at 60°F
  - Pressure Recorder Serial 94-149142, two pen
  - Regulator Canadian Meter Co. Reliance, Type 1813B  
Serial 100-149146, Inlet pressure 60 psig, Internal relief set at 3 psig.
  
3. Versiculite Ltd., 1775 52nd Ave., Lachine, P.Q. X
  - Meter Type AL-425, Serial 12-149087 (CNC 51816)  
Compensated at 2.14 psi, compensated at 60°F



(cont'd)

3. - Pressure recorder Serial 94-149143, two pen  
- Regulator Canadian Meter Co. Reliance, Type 1213B  
Serial 100-149147, Inlet Pressure 60 psig, Internal Relief  
set at 3 psig.
  
4. - Diamond-Recall Inc., 9205 Fairway Ave., Lucerne, P.Q. X  
- Meter Type 2225, Serial 12-149228 (SNC 51215)  
Compensated at 2.14 psig, Compensated at 60<sup>0</sup>F  
- Pressure recorder Serial 94-149144, two pen  
- Regulator Canadian Meter Co. Reliance, Type 1213B  
Serial 100-149148, Inlet Pressure 60 psig, Internal  
Relief set at 3 psig.

Approval of these Pressure Compensated Indexes is temporary  
pending completion of approval tests and finalization of related requirements.

*[Handwritten signature]*  
W. J. S. Fraser,  
Chief, Electricity and Gas Division  
Standards Branch.

c.c. Mr. J. S. Dumas, District Inspector of E&G, Montreal  
Canadian Meter Company, Milton, Ontario.  
Regional Office - Montreal

